



Never stop looking®

## CE DECLARATION OF CONFORMITY

**iCAD, Inc.**  
**98 Spit Brook Road, Suite 100**  
**Nashua, New Hampshire 03062, USA**  
**+ 1 937-431-1464**  
**Fax: +1-603-880-3843**

Declares under sole responsibility that the product **PowerLook AMP™ (Product No: DSC010-01-EU, DSC010-03-EU, DSC010-04-EU, DSC010-05-EU, DSC010-06-EU-A, DSC010-07-EU, DSC010-08-EU, DSC010-09-EU, D70088, D70089, D70090, D70091, D70092, D70093, D70094, D70096, D70097, D70098, D70099, D70084, D70085, DSC010-01-EU-A, DSC010-03-EU-A, DSC010-04-EU-A, DSC010-05-EU-A, DSC010-06-EU-A, DSC010-07-EU-A, DSC010-08-EU-A, DSC010-09-EU-A, D4D0146, and D4D0147)** to which this declaration relates meets the essential health and safety requirements and is in conformance with the relevant EC directives listed below using the relevant section of the EC standards and other normative documents:

IEC 60950-1  
EN 55022  
EN 55024  
Medical Device Directive 93/42/EEC

**MDD Class:** Class I. The classification is based on the requirements of Rule 12 of Annex IX, of the Medical Device Directive.

**Quality System Certification:** ISO 13485:2003 as indicated on certificate number 9213, granted by Intertek.

**CE Mark:** Article 17 of the Medical Device Directive 93/42/EEC.

**Date:** August 15, 2012

iCAD, Inc., Authorized Representative in the European Community (as defined in Article 14 of the medical Device Directive: 93/42/EEC):

MDSS GmbH  
Schiffgraben 41  
30175 Hannover, Germany

Management Representative, John A. DeLucia  
Vice President of Regulatory Affairs, Clinical Affairs and Quality Assurance